
EN55024 IMMUNITY TEST REPORT

Applicant: Orbital Therapy, LLC.
8 Alfred Circle
Bedford, MA 01730

Equipment Under Test: Fiber-Optic Dosimeter

Models covered by Testing: HRD-6 operated on USB power from a pc

Test Location: IQS, a Division of Degree Controls
257 Simarano Drive
Marlboro, MA 01752-1298

Test Dates: January 15 and February 3, 2009

Report Date: February 09, 2009

EN 61000-4-2:1995, A1:1998, A2:2001 (Electrostatic Discharge) - Passed
EN 61000-4-3:2006 (Radiated Immunity) - Passed
EN 61000-4-4:1995, A1:2001, A2:2002; 2004 (Electrical Fast Transients/Burst) – Passed

This product does comply with the requirements of EN 55024:1998, A1:2001, A2:2003.
Please refer to attached individual reports for detailed testing results.

Reviewed by: _____



Charles Nasser

This report has been prepared to record the Immunity test results of the product tested, as identified above and may only be used to document the testing as identified herein. This report may only be reproduced in full. Any distribution of the information contained in this document requires the prior approval from Orbital Therapy, LLC.
