
EMISSION TEST REPORT

Applicant: Orbital Therapy, LLC.
8 Alfred Circle
Bedford, MA 01730

Equipment Under Test: Fiber-Optic Dosimeter

Models covered by Testing: HRD-6 operated on USB power from a pc

Test Location: IQS, a Division of Degree Controls
257 Simarano Drive
Marlboro, MA 01752-1298

Test Date: January 16, 2009

Report Date: February 09, 2009

CFR47 Part 15: 8/14/2006	(FCC) - Passed
EN55022: 2006	(EU Emissions) - Passed
CISPR22: 3 rd edition 2006	(Worldwide Emissions) - Passed
AS/NZS: CISPR 22: 3 rd edition 2006	(Australia & New Zealand Emissions) - Passed
ICES-003: issue 4, 2/2004	(Canadian Emissions) - Passed

Please refer to attached individual reports for detailed testing results.

Reviewed by: _____



Charles Nasser

This report has been prepared to record the Emissions test results of the product tested, as identified above and may only be used to document the testing as identified herein. This report may only be reproduced in full. Any distribution of the information contained in this document requires the prior approval from Orbital Therapy, LLC.
